

アプリケーションノート

# Pseudoephedrine HCL and Triprolidine HCL

Waters Corporation



This is an Application Brief and does not contain a detailed Experimental section.


## Abstract

This application brief highlights the analysis of Pseudoephedrine HCL and Triprolidine HCL using Symmetry Columns.


## Introduction

The compounds analyzed in this study are:

- 1. Pseudoephedrine HCL
- 2. Triprolidine HCL



1. Pseudoephedrine HCL



2. Triprolidine HCL

# Experimental

#### **HPLC** Method

Column:	Symmetry C <sub>8</sub> , 3.9 x 150 mm, 5 $\mu$ m
Guard column:	Symmetry Guard Column 3.9 x 20 mm, 5 μm
Part numbers:	Column - WAT046970, Guard - WAT054250
Mobile phase A:	50 mM potassium phosphate, pH 3.0

Mobile phase B:	Acetonitrile
Flow rate:	1.0 mL/min
Injection volume:	5 μL of 2.88 μg/mL pseudoephedrine and 120 μ g/mL triprolidine extracted tablet sample
Detection:	UV @ 261 nm

### Gradient Table


Time	Profile	
(min)	%A	%B
0	85	15
1	85	15
15	50	50

## USP Tailing Factors

1. 1.53

2.1.16

# Results and Discussion



# Featured Products

© 2021 Waters Corporation. All Rights Reserved.