



# Pseudoephedrine HCL and Triprolidine HCL

Waters Corporation



This is an Application Brief and does not contain a detailed Experimental section.

### **Abstract**

This application brief highlights the analysis of Pseudoephedrine HCL and Triprolidine HCL using Symmetry Columns.

### Introduction

The compounds analyzed in this study are:

- 1. Pseudoephedrine HCL
- 2. Triprolidine HCL

#### 1. Pseudoephedrine HCL

$$CH_{3} = CH_{2} - N$$

$$C = CH_{2} - N$$

$$HC1 \cdot H_{2}C$$

#### 2. Triprolidine HCL

### Experimental

#### **HPLC** Method

Column: Symmetry  $C_8$ , 3.9 x 150 mm, 5  $\mu m$ 

Guard column: Symmetry Guard Column 3.9 x 20 mm, 5  $\mu$ m

Part numbers: Column - WAT046970, Guard - WAT054250

Mobile phase A: 50 mM potassium phosphate, pH 3.0

Mobile phase B: Acetonitrile

Flow rate: 1.0 mL/min

Injection volume: 5  $\mu$ L of 2.88  $\mu$ g/mL pseudoephedrine and 120  $\mu$ 

g/mL triprolidine extracted tablet sample

Detection: UV @ 261 nm

### **Gradient Table**


Time	Profile	
(min)	%A	%B
0	85	15
1	85	15
15	50	50

### **USP Tailing Factors**

1. 1.53

2. 1.16

## Results and Discussion



Featured Products

WA31763.138, June 2003	
$\wedge$	
© 2021 Waters Corporation. All Rights Reserved.	