Waters™

Note d'application

Terfenadine and Related Compound

Waters Corporation



This is an Application Brief and does not contain a detailed Experimental section.

Abstract

This application brief highlights the analysis of Terfenadine and related compounds using Symmetry C_8 Columns.

Introduction

The compounds analyzed in this study are:

- 1. Terfenadine
- 2. Terfenadine Related Compound ARS

1. Terfenadine

Experimental


HPLC Method

Column: Symmetry C_8 , 4.6 x 250 mm, 5 μm

Part number: WAT054270

| Mobile phase: | 100 mM triethylammonium phosphate, pH 7.0/methanol 20:80 |
|--------------------|--|
| Flow rate: | 1.0 mL/min |
| Injection volume: | A. 15 μ L of mixture with 0.6 mg/mL terfenadine and 0.15 mg/mL of terfenadine related compound ARS B. 15 μ L of 8mg/mL crushed tablet (equivalent to 0.6mg/mL terfenadine) |
| Detection: | UV @ 235 nm |
| USP Tailing Factor | |
| 1. 1.0 | |

Results and Discussion



© 2021 Waters Corporation. All Rights Reserved.