# Waters™



## Lidocaine HCL

Waters Corporation

This is an Application Brief and does not contain a detailed Experimental section.

#### **Abstract**

This application brief demonstrates the analysis of Lidocaine HCL using Symmetry Columns.

### Introduction

The compound analyzed in this study is Lidocaine HCL.

# 1. Lidocaine HCL


# 2. Methylparaben

## Experimental

#### **HPLC** Method

Column: Symmetry Shield RP  $_{18},\,4.6$  x 250 mm, 5  $\mu m$ Part number: 186000112 Mobile phase: 50 mL glacial acetic acid in 930 mL of water, pH adjusted to 3.4 with 0.1 N sodium hydroxide/70% buffer/30% acetonitrile Flow rate: 1.5 mL/min Injection volume: 20 μL Detection: UV @ 254 nm **USP Tailing Factors** 1. 2.49 2. 0.99

### Results and Discussion



## Featured Products

WA31763.95, June 2003

© 2022 Waters Corporation. All Rights Reserved.