

Nota de aplicación

## Analysis of Estradiol USP Method Using XBridge C<sub>18</sub>

---

Waters Corporation



This is an Application Brief and does not contain a detailed Experimental section.

---


### Abstract

This application brief highlights the analysis of estradiol USP method using XBridge C<sub>18</sub> Columns.


---

# Introduction


## Compounds



### 1. Ethylparaben



### 2. Estradiol



### 3. Estrone

---

# Experimental

## Test Conditions

Column:

XBridge C<sub>18</sub>, 5 μm, 4.6 x 150 mm p/n: 186003116


Isocratic Mobile Phase:

water/acetonitrile (45:55)

Flow Rate: 1.0 mL/min  
Injection Volume: 10 µL  
Sample Concentration: 20 µg/mL in water/ACN (50:50)  
Temperature: 25 °C  
Detection: UV @ 205 nm  
Instrument: Alliance 2695 with 2996 PDA

---

## Results and Discussion



1. Ethylparaben 2. Estradiol 3. Estrone.

---

## Featured Products

Alliance HPLC <<https://www.waters.com/514248>>

WA60164, April 2007

© 2021 Waters Corporation. All Rights Reserved.