

Note d'application

## Use of Nitrogen as Carrier Gas for GC-MS/MS With Atmospheric Pressure Ionization (APGC) for the Determination of Pesticide Residues in Food

---

David Gould, Simon Hird, Frank Dorman, Douglas Stevens

Waters Corporation

---

### Abstract

Gas chromatography coupled to Tandem Quadrupole Mass Spectrometry (GC-MS/MS) is a powerful analytical technique used for the detection and quantification of pesticide residues in food and environmental samples. This technique historically utilizes helium as a carrier gas. In recent year there have been difficulties sourcing helium worldwide, leading to significant increases in price and difficulties in availability. This has created an increased level of interest for alternative carrier gases for gas chromatography in general. Nitrogen is readily available, relatively inexpensive, and safe compared to other options. The following application demonstrates the ease of transferring GC-MS/MS methods, utilizing an atmospheric pressure ionization source (APGC™ technology) using helium carrier gas to nitrogen carrier gas. APGC achieves improved selectivity and sensitivity when compared to electron ionization (EI) analysis, with APGC demonstrating equivalent performance when using helium or nitrogen as carrier.

A routine method of over 200 pesticides was evaluated on the APGC Xevo™ TQ-S micro System using both carrier gases was performed on a cucumber extract Sample preparation utilized a modified Quick, Easy, Cheap,

---

Effective, Rugged, and Safe (QuEChERS) method. Extracts of infant food (cottage pie) were also prepared for pesticide analysis using nitrogen as the carrier to demonstrate equivalent performance of each gas when using APGC, simply by using a smaller diameter column and scaling the analytical method gas flow parameters.

## Benefits

Proving equivalent chromatographic separation and performance when switching to nitrogen carrier with APGC can introduce a viable alternative for GC experiments for labs that are under pressure caused by helium's cost and scarcity.

---

## Introduction

Helium is the most commonly used carrier gas for Gas Chromatography.<sup>1</sup> Helium's usage is based on historical applications and instrumentation design, and it was generally assumed to be the preferred carrier gas when its use was not as restricted and/or costly as it is today. Helium is a byproduct of natural gas extraction, and certain helium-producing regions have experienced political instability, resulting in disruptions to the global supply chain. Oil and gas production fluctuations also impact the availability of helium as it primarily sourced as a by-product of these processes. Helium is also used for other technologies such as magnetic resonance imaging (MRI) machines, semiconductor manufacturing, and superconductivity research, further increasing demand. Additionally, some of these applications areas, most notably medical diagnosis, are prioritized over chromatographic applications. This limited supply-demand balance has led to price escalation.<sup>2</sup>

The advantages of using APGC-MS/MS for pesticide analysis can be found here:

<https://www.waters.com/webassets/cms/library/docs/720007815en.pdf> <

<https://www.waters.com/webassets/cms/library/docs/720007815en.pdf>>

<https://www.waters.com/content/dam/waters/en/app-notes/2022/720007708/720007708-en.pdf> <

<https://www.waters.com/content/dam/waters/en/app-notes/2022/720007708/720007708-en.pdf>>

<https://www.waters.com/content/dam/waters/en/app-notes/2022/720007654/720007654-en.pdf> <

<https://www.waters.com/content/dam/waters/en/app-notes/2022/720007654/720007654-en.pdf>>

Helium (He) is becoming a costly choice for a carrier gas for gas chromatography-mass spectrometry (GC-MS)

as supplies have been restricted resulting in shortages. Chromatographers have found that He is much more expensive and is not always available when needed. Many have considered switching from He to hydrogen as an alternative. However, there are perceived concerns about its safety, reactivity on older instruments, reduced performance compared to He, and the length and cost of the transition process. This all combines to make the case for using nitrogen. Nitrogen is an inert, renewable gas which is relatively inexpensive compared to the alternatives. Nitrogen already has its use in non-MS detectors but is met with trepidation in MS laboratories due to vacuum source flow requirements. With APGC there is no need for a high vacuum and ionization takes place around a corona pin which induces a nitrogen plasma, so using nitrogen as a carrier gas will have no adverse effects. The mechanism of the APGC technique is gas-phase ion molecule reactions by charge or proton transfer. For pesticide residue analysis, source conditions are adjusted to have just enough water present in the ionization chamber to give sufficient sensitivity for compounds that prefer either charge or proton transfer. The instrument setup has very dry conditions, achieved by using appropriate quality nitrogen gas and filtration. Water is then added vials in a holding tray which is placed within the source enclosure.

The objective of this study is to demonstrate the ease of which one can switch to using nitrogen as carrier gas for APGC-MS/MS without having to sacrifice on performance.

---

## Experimental

### Sample Description

A QuEChERS extraction was performed on samples of cucumber and infant food (cottage pie) to make matrices in acetonitrile for bracketed calibrations of 203 GC-amenable pesticides (RESTEK p/n: 32562). Both matrices were extracted using a modified version of the QuEChERS CEN method.<sup>4</sup> For infant food this also involves SPE clean up by passing through an Oasis™ PRIME HLB cartridge. Matrix matched standards were made over the range of 0.00025 to 0.02 mg/kg and analyzed using the APGC Xevo TQ-S micro System, changing from helium to nitrogen carrier on multiple occasions.

### Method Conditions

The changeover process to nitrogen carrier is simply a case of switching the gas supply over and with the appropriate filters even a standard pressurized cylinder of nitrogen can be used. Nitrogen has a lower optimum linear velocity (10–15 cm/s) compared to helium (25–35 cm/s) thus a lower column flow must be used for

nitrogen carrier. Therefore, in order to achieve equivalent retention times, the GC method needs to be translated to a smaller length column with a matching phase ratio as the column used for helium analysis. There are tools available online to help with these transitions, the example shown in Figure 1 was used for this analysis .

**EZGC Method Translator**

Carrier Gas	Original	Translation
	Helium	Nitrogen

**Column**

Parameter	Original	Translation	Unit
Length	30.00	20.00	m
Inner Diameter	0.25	0.15	mm
Film Thickness	0.25	0.15	µm
Phase Ratio	250	250	

**Control Parameters**

Parameter	Original	Translation	Unit
Column Flow	2.00	0.50	mL/min
Average Velocity	43.63	28.65	cm/sec
Holdup Time	1.15	1.16	min
Inlet Pressure	23.62	26.26	psi
Outlet Pressure (abs)	14.70	14.70	psi

**Oven Program**

Isothermal

Ramps

Ramp	Ramp Rate (°C/min)	Temp (°C)	Hold Time (min)	Ramp Rate (°C/min)	Temp (°C)	Hold Time (min)
1	8.5	330	5	8.3	330	5.1

**Control Method**

Constant Flow

**Results** Solve for  Efficiency  Speed  Translate  Custom

Run Time	34.24	35.02	min
Speed		0.98	x

Use Flow Calculator Values

Figure 1. EZGC Method Translator.<sup>5</sup>

By entering the method parameters for helium the translator will try to match the 'Holdup Time' for the nitrogen method and will give a reasonable starting point for a new method, a little method development work may be required with the column flow to match retention times exactly. After initial method development, a solvent

focusing step was added to the oven ramp for the nitrogen method for some of early eluting compounds.

## GC Conditions

Parameter	Helium	Nitrogen
Column	RESTEK SVOCms 30 m x 0.25 mm ID x 0.25 $\mu$ m	RESTEK SVOCms 20 m x 0.15 mm ID x 0.15 $\mu$ m
Oven program	90 °C for 1 min Ramp to 330 °C at 8.5 °C/min Hold for 5 min	75 °C for 0.7 min Ramp to 106.6 °C at 15.8 °C/min, no hold Ramp to 330 °C at 8.3 °C/min Hold for 5.1 min
Gas flow rate:	2 mL/min (constant flow mode)	0.38 mL/min (constant flow mode) – cucumber matrix 0.45 mL/min (constant flow mode) – infant food matrix
Injection type:	Pulsed splitless	Pulsed splitless
Inlet temp:	250 °C	250 °C
Pulse time:	1.2 min	1.2 min
Pulse pressure:	32 psi	50 psi
Purge flow:	30 mL/min	30 mL/min
Septum purge flow:	3 mL/min	3 mL/min
Inlet liner:	Restek Topaz 4.0 mm ID Single Taper Inlet Liner	Restek Topaz 4.0 mm ID Single Taper Inlet Liner
Injection volume:	1 $\mu$ L	1 $\mu$ L
Makeup gas:	Nitrogen at 350 mL/min	Nitrogen at 350 mL/min
Transfer line temp:	280 °C	280 °C

---

## Results and Discussion

### Method Translation

Due to nitrogen having a lower optimum linear velocity than helium, the gas flow rate is much lower for the nitrogen method, so a smaller column length column is needed otherwise retention times could be twice as long. Matching the phase ratio of the column to the one used for helium analysis ensures chromatographic performance is retained.

The switch of carrier gas was an easy transition as scaling the column dimensions along with a small amount of method development allows for the retention times to be matched from one carrier gas to another. This allowed many of the method files used for data processing to be kept the same for each type of carrier gas. The nitrogen analysis showed equivalent calibration coefficients for the historically “difficult” compounds as well as equivalent resolution of critical pairs. Both carrier gases analyzed on APGC Xevo TQ-S micro show superior sensitivity to an EI based system and demonstrated consistent performance when using either helium or nitrogen as carrier.

---

## Chromatographic Performance

Figure 2 shows a comparison of early eluting peaks in the infant food matrix. Using the parameters calculated in the method translator it can be shown that peaks will lie within the  $\pm 0.2$  min range that means the same processing file can be used.




Figure 2. Early eluting compounds in infant food.

Figure 3 shows late eluting peaks from the cucumber matrix, given more time to develop the method and experiment with column flow, the retention times matched up perfectly. This allowed for the same processing method to be used for each carrier gas across both matrices.




Figure 3. Late eluting compounds in cucumber.

Figure 4 shows the chromatograms for the critical pair, profluralin (red), and terbutylazine (purple) where nitrogen carrier offers superior resolution. This could be due to the lower linear velocity of nitrogen giving each analyte slightly more time to interact with the stationary phase of the column.




Figure 4. Critical pair resolution in helium (left) and nitrogen (right).

## Repeatability

Each of the 6 internal standards used in this analysis returned residuals under 10% over 12 injections for each matrix. Figure 5 shows the comparison of an internal standard helium (left) and nitrogen (right) over multiple injections illustrating the residual plot with all injection falling within a 10% error range.




Figure 5. Internal standard repeat injections for helium (left) and nitrogen (right).

## Sensitivity & Calibration Characteristics

The graph shown in Figure 6 shows the ratio of the helium response to the nitrogen response for the 0.01 mg/kg standard in infant food. This demonstrates the parity in sensitivity between the two carrier gases as approximately 59% of the analytes have a ratio between 0.5 & 1.5. 85% of the analytes give ratios less than 2.






Figure 6. Response for helium divided by response for nitrogen for 0.01 mg/kg standard.

What follows are a list of historically "difficult" compounds detected at maximum residue level (MRL) 0.01 mg/kg all analyzed with nitrogen carrier, along with a comparison in sensitivity for cypermethrin and captan over each carrier gas. They all display residual plots within 20%, conforming to SANTE guidelines and  $R^2$  coefficients of  $>0.99$ .




Figure 7. Chromatogram comparison for cypermethrin in infant food at 0.01 mg/kg.




Figure 8. Chromatogram comparison for captan in infant food at 0.01 mg/kg.




Figure 9. Bracketed calibration graphs for cypermethrin and captan residues in infant food.




Figure 10. Chromatograms for a selection of pesticide residues in infant food at 0.01 mg/kg.




Figure 11. Bracketed calibration graphs for a selection of pesticide residues in infant food.

## Conclusion

This application note describes the easy conversion of an existing GC method for the determination of pesticide residues from He to N<sub>2</sub> as a carrier gas. As the ion source is not constrained by vacuum, the APGC system is able to switch to using N<sub>2</sub> as a carrier gas without a loss in performance and can be returned to He at the turn of a valve. After scaling the column dimensions and adjustment of parameters, separations and run time were almost identical with no impact upon sensitivity. APGC with N<sub>2</sub> as a carrier gas has been shown to be reliable for the determination of pesticide residues in food and offers considerable saving on carrier gas costs with no supply issues

---

## References

1. <https://www.peakscientific.com/discover/news/how-is-helium-gas-obtained/> <  
<https://www.peakscientific.com/discover/news/how-is-helium-gas-obtained/>> .
2. <https://www.innovationnewsnetwork.com/helium-shortage-4-0-what-caused-it-and-when-will-it-end/29255/> <<https://www.innovationnewsnetwork.com/helium-shortage-4-0-what-caused-it-and-when-will-it-end/29255/>> .
3. European Committee for Standardisation (CEN) EN 15662:2018. Foods of Plant Origin - Multimethod for the Determination of Pesticide Residues Using GC- and LC- Based Analysis Following Acetonitrile Extraction/Partitioning and Clean-up by Dispersive Spe - Modular Quechers-Method.
4. <https://ez.restek.com/ezgc-mtfc> <<https://ez.restek.com/ezgc-mtfc>> .

---

## Featured Products

[Waters Atmospheric Pressure Gas Chromatography \(APGC\)](https://www.waters.com/10100362) <<https://www.waters.com/10100362>>

[Xevo TQ-S micro Triple Quadrupole Mass Spectrometry](https://www.waters.com/134798856) <<https://www.waters.com/134798856>>

[MassLynx MS Software](https://www.waters.com/513662) <<https://www.waters.com/513662>>

[TargetLynx](https://www.waters.com/513791) <<https://www.waters.com/513791>>

720008044, September 2023



© 2023 Waters Corporation. All Rights Reserved.

[Conditions d'utilisation](#) [Confidentialité](#) [Marques](#) [Carrières](#) [Cookies](#) [Préférences de cookies](#)