

Application Note

Simple and Selective Bioanalytical Plasma Sample Extraction Using Oasis® MCX SPE for High Analyte Recovery

Mary Trudeau, Margot Lee

Waters Corporation

Este é um Resumo de aplicações e, por isso, não inclui uma seção de experimento detalhada.

Abstract

This application demonstrates the simple and selective bioanalytical plasma sample extraction using mixed-mode ion-exchange solid phase extraction (SPE), requiring no protocol method development, for high extraction MCX SPE recovery of several tyrosine kinase inhibitors (TKI) used in the treatment of leukemia.

Experimental




Figure 1. Demonstration of Oasis MCX SPE bioanalytical extraction performance of several TKI pharmaceuticals, requiring no optimization of the starting protocol, resulting in > 72.6% recovery and matrix effects between -3.8 and -11.0% from plasma.

Results and Discussion

High analyte recovery and selectivity from plasma using Oasis MCX SPE for top-selling small molecule tyrosine kinase inhibitor pharmaceuticals in 2023.¹

References

1. Williams, RE and Leatherwood HM Top 200 Small Molecule Drugs by Retail Sales in 2023.<https://bpb-us-e2.wpmucdn.com/sites.arizona.edu/dist/9/130/files/2024/05/2023Top200SmallMoleculePosterV5.pdf> <<https://bpb-us-e2.wpmucdn.com/sites.arizona.edu/dist/9/130/files/2024/05/2023Top200SmallMoleculePosterV5.pdf>> (accessed 21 June 2024).

Featured Products

720008424, July 2024



© 2024 Waters Corporation. All Rights Reserved.

[Termos de Uso](#) [Política de Privacidade](#) [Marcas comerciais](#) [Carreiras](#) [Avisos jurídicos e de privacidade](#) [Cookies](#) [Preferências de cookies](#)