Waters™

Note d'application

Extraction of Oligonucleotides From Plasma Samples Across Multiple Species Using OligoWorks[™] SPE Microplate Kit

Mary Trudeau, Matthew A. Lauber

Waters Corporation

Ce document est une note d'application et ne contient pas de section détaillée concernant l'expérimentation.

Abstract

Oligonucleotide therapeutics are a key focus area for many drug developers today given their powerful ability to address disease biology at the level of gene transcription and translation with high target specificity and low toxicity. They are designed with specific types of modifications that enhance protein binding, half-life, and cellular uptake. Given their diversity, developing robust bioanalytical methods for oligonucleotide analysis can be quite challenging, and sample preparation is often its most challenging aspect due to the tight oligonucleotide-to-protein binding that occurs in biomatrices, often leading to low extraction recoveries and poor sensitivity. This work demonstrates use of the OligoWorks SPE Microplate Kit with RapiZyme™ Proteinase K Digestion to effectively extract oligonucleotides from plasma samples across multiple preclinical animal species including human. Using the starting protocol with no adjustments, oligonucleotide recoveries between 56–115% were obtained.

Extraction of Oligonucleotides From Plasma Samples Across Multiple Species Using OligoWorks[™] SPE Microplate 1 Kit

Experimental

Sample Preparation and Extraction

OligoWorks sample preparation and SPE extraction protocol

RapiZyme Proteinase K digestion sample pretreatment



Figure 1. Graphical representation of the OligoWorks SPE Microplate Kit (p/n: 186010614) sample preparation protocol, with plasma sample digestion pretreatment using RapiZyme Proteinase K, followed by WAX SPE using the OligoWorks SPE microplate. LC-MS analysis of plasma sample extracts was performed with an ACQUITYTM UPLCTM I-Class PLUS (FL) coupled to a Waters XevoTM TQ-XS Tandem Quadrupole Mass Spectrometer using multiple-reaction monitoring mode (MRM). Chromatographic separation was achieved ACQUITY UPLC Premier Oligonucleotide C₁₈ Column, 130 Å, 1.7 µm, 2.1 x 50 mm (p/n: 186009484).

Results




Figure 2. OligoWorks SPE Microplate Kit extraction performance (no internal standard correction) demonstrating high plasma* recoveries (1-hour RapiZyme Proteinase K Digestion, 55 °C), for a diversity of oligonucleotides. *A 1:1 water dilution of digested plasma sample was applied prior to SPE loading. The 1:1 dilution minimized oligonucleotide loss (break-through) on SPE sample load, ensuring high SPE recovery.

Ordering Information

Description	P/N
OligoWorks SPE Microplate Kit	186010614
ACQUITY Premier Oligonucleotide C ₁₈ Column, 130Å, 1.7 μm 2.1 × 50 mm	186009484
QuanRecovery™ with MaxPeak, 700 µL plate	186009184
Polypropylene cap mat round well for 96-well	186009452

Featured Products

ACQUITY UPLC I-Class PLUS System <

Extraction of Oligonucleotides From Plasma Samples Across Multiple Species Using OligoWorks™ SPE Microplate 3 Kit https://www.waters.com/nextgen/global/products/chromatography/chromatography-systems/acquity-uplc-iclass-plus-system.html>

Xevo TQ-XS Triple Quadrupole Mass Spectrometer <https://www.waters.com/nextgen/global/products/massspectrometry/mass-spectrometry-systems/xevo-tq-xs.html>

RapiZyme Proteinase K Digestion Module https://www.waters.com/nextgen/global/products/standards-and-reagents/rapizyme-proteinase-k-digestion-module.html

720008499, September 2024

 \wedge

© 2024 Waters Corporation. All Rights Reserved.

Conditions d'utilisation Politique de confidentialité Marques Carrières Mentions légales et déclaration de confidentialité Cookies Préférences de cookies